

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 04 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130173
Name		J. Heinbaugh	ID# 5828 Date 12-07-09
A	Agency	AST	Phone # 800-3263
Instrument Location		Glennallen	
Alco S/N		0124625	Target Value 0.083 High Pressure 1000
B	Alco Test Values	0.079	0.079
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		JLB	CBS 1/9/10
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of January, 2010.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130173

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

DECEMBER 27, 2009

OPERATOR'S NAME:

WENBAUGH

OPERATOR'S NUMBER: 5828

SUBJECT'S LAST NAME:

SUBJECT'S FIRST NAME/MI :

CALL #: UOC

DEPT AGENCY: GNK0

OPER REPORT: UOC

TEST TYPE: U

ALCOHOL TARGET VALUE: 0.083

ALCO S/N: X124625

--- BREATH ANALYSIS ---

ALCOHOL DETECTED: 0.083 in

ALCOHOL TEST	.079	12:43
FLAME TEST	.000	12:44
INTERNAL STANDARD	VERIFIED	12:44
ALCOHOL IN 20.62 in	.079	12:44
FLAME TEST	.000	12:45
SUBJECT SAMPLE	.000	12:46
FLAME TEST	.000	12:46
ALCOHOL IN 20.62 in	.079	12:47
FLAME TEST	.000	12:48

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130173

DECEMBER 27, 2009

TIME 12:41

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 28.60 in

FLAME DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

LEVITATOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

1 123456789:;=>?@ABCDEF6  
PQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
PQRSTUVWXYZ{|}~